

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120324\
 Data File : VV038405.D
 Acq On : 03 Dec 2024 14:01
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5304

Quant Time: Dec 03 21:48:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120324.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 03 21:43:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	8101	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	8116	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	3810	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	4085	0.546	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	110.000%
4) 1,2-Dichloroethane-d4	4.969	65	3370	0.587	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	118.000%
7) 1,2-Dichloropropane-d6	5.995	67	4129	0.543	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	108.000%
8) Toluene-d8	7.253	98	8926	0.501	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3396	0.575	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	116.000%
Target Compounds						
						Qvalue
3) Vinyl chloride	1.280	62	5057	0.527	ug/L	98
6) Trichloroethene	5.852	95	3422	0.497	ug/L	98
9) 1,2-Dibromoethane	8.305	107	2363	0.532	ug/L	99
12) 1,2,3-Trichloropropane	10.197	75	2103	0.530	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.357	75	369	0.542	ug/L #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120324\
Data File : VV038405.D
Acq On : 03 Dec 2024 14:01
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5304

Quant Time: Dec 03 21:48:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120324.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 03 21:43:44 2024
Response via : Initial Calibration

