

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120325\
 Data File : VV039451.D
 Acq On : 03 Dec 2025 12:56
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV275

Quant Time: Dec 04 01:09:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Dec 04 00:58:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	7813	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.785	117	7688	0.500	ug/L	0.00
12) 1,4-Dichlorobenzene-d4	11.175	152	4312	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.295	65	3791	0.483	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	96.000%		
4) 1,2-Dichloroethane-d4	4.970	65	2285	0.439	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	88.000%		
7) 1,2-Dichloropropane-d6	6.020	67	3324	0.512	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	102.000%		
8) Toluene-d8	7.238	98	5292	0.484	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	96.000%		
10) 1,1,2,2-Tetrachloroeth...	10.133	84	1798	0.492	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	98.000%		
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.295	62	7743	0.458	ug/L	100
6) Trichloroethene	5.853	95	4974	0.522	ug/L	98
9) 1,2-Dibromoethane	8.290	107	2895	0.460	ug/L	100
11) 1,1,2,2-Tetrachloroethane	10.155	83	3795	0.492	ug/L	99
13) 1,2,3-Trichloropropane	10.200	75	2193	0.483	ug/L	97
14) 1,2-Dibromo-3-chloropr...	12.351	75	438	0.477	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120325\
Data File : VV039451.D
Acq On : 03 Dec 2025 12:56
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV275

Quant Time: Dec 04 01:09:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Dec 04 00:58:46 2025
Response via : Initial Calibration

