

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120325\
 Data File : VV039449.D
 Acq On : 03 Dec 2025 11:52
 Operator : SY/MD
 Sample : VSTD00173
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD001273

Quant Time: Dec 04 00:44:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Dec 04 00:43:25 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Difluorobenzene	5.566	114	7477	0.500 ug/L	0.00
5) Chlorobenzene-d5	8.785	117	7812	0.500 ug/L	0.00
12) 1,4-Dichlorobenzene-d4	11.175	152	4678	0.500 ug/L	0.00
System Monitoring Compounds					
2) Vinyl Chloride-d3	1.295	65	7342	0.976 ug/L	0.00
4) 1,2-Dichloroethane-d4	4.970	65	4762	0.956 ug/L	0.00
7) 1,2-Dichloropropane-d6	5.996	67	6633	1.006 ug/L	-0.02
8) Toluene-d8	7.238	98	11524	1.032 ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.133	84	3803	1.024 ug/L	0.00
Target Compounds					
					Qvalue
3) Vinyl chloride	1.295	62	14821	0.916 ug/L	98
6) Trichloroethene	5.829	95	9481	1.001 ug/L	91
9) 1,2-Dibromoethane	8.274	107	5986	0.936 ug/L #	96
11) 1,1,2,2-Tetrachloroethane	10.155	83	7698	0.982 ug/L	98
13) 1,2,3-Trichloropropane	10.177	75	4739	0.990 ug/L	94
14) 1,2-Dibromo-3-chloropr...	12.351	75	872	0.875 ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120325\
Data File : VV039449.D
Acq On : 03 Dec 2025 11:52
Operator : SY/MD
Sample : VSTD00173
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD001273

Quant Time: Dec 04 00:44:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Dec 04 00:43:25 2025
Response via : Initial Calibration

