

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV120420\
 Data File : VV019477.D
 Acq On : 04 Dec 2020 14:39
 Operator : SY/MD
 Sample : L4955-01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG4C7

Quant Time: Dec 05 02:54:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM112320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 05 02:40:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	312167	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	315151	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	145109	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	98151	31.48	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	62.96%
7) Chloroethane-d5	1.58	69	94297	37.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	74.32%
11) 1,1-Dichloroethene-d2	2.12	63	172914	31.16	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.32%
21) 2-Butanone-d5	3.89	46	191662	100.86	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.86%
24) Chloroform-d	4.37	84	233434	46.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.62%
26) 1,2-Dichloroethane-d4	5.05	65	167477	50.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.52%
32) Benzene-d6	5.07	84	449676	48.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.52%
36) 1,2-Dichloropropane-d6	6.09	67	157230	50.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.44%
41) Toluene-d8	7.33	98	381729	45.73	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.46%
43) trans-1,3-Dichloropropene-	7.64	79	69669	46.52	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.04%
47) 2-Hexanone-d5	8.10	63	91431	74.85	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	74.85%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	162826	40.57	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.14%
64) 1,2-Dichlorobenzene-d4	11.65	152	143489	50.92	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.84%

Target Compounds					Ovalue
13) Acetone	2.17	43	5767	3.969 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV120420\
Data File : VV019477.D
Acq On : 04 Dec 2020 14:39
Operator : SY/MD
Sample : L4955-01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG4C7

Quant Time: Dec 05 02:54:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM112320WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 05 02:40:32 2020
Response via : Initial Calibration

