

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV120420\
 Data File : VV019481.D
 Acq On : 04 Dec 2020 16:21
 Operator : SY/MD
 Sample : VIBLK83
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK83

Quant Time: Dec 05 03:13:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM112320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 05 02:40:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	287806	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	289909	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	131612	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	148010	51.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	102.98%
7) Chloroethane-d5	1.58	69	121902	52.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	104.22%
11) 1,1-Dichloroethene-d2	2.12	63	217363	42.48	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.96%
21) 2-Butanone-d5	3.90	46	178076	101.64	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.64%
24) Chloroform-d	4.37	84	257729	55.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.92%
26) 1,2-Dichloroethane-d4	5.05	65	184771	60.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	121.50%
32) Benzene-d6	5.07	84	502693	58.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	117.30%
36) 1,2-Dichloropropane-d6	6.09	67	173930	60.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	121.98%#
41) Toluene-d8	7.33	98	425219	55.38	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.76%
43) trans-1,3-Dichloropropene-	7.64	79	75145	54.55	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	109.10%
47) 2-Hexanone-d5	8.11	63	92774	82.56	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.56%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	187066	50.67	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.34%
64) 1,2-Dichlorobenzene-d4	11.65	152	152089	59.51	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	119.02%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV120420\
Data File : VV019481.D
Acq On : 04 Dec 2020 16:21
Operator : SY/MD
Sample : VIBLK83
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK83

Quant Time: Dec 05 03:13:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM112320WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 05 02:40:32 2020
Response via : Initial Calibration

