

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120423\
 Data File : VV033089.D
 Acq On : 04 Dec 2023 10:08
 Operator : SY/MD
 Sample : VV1204MBL01
 Misc : 5.0G/5.0mL/100uL/5mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK230

Quant Time: Dec 05 03:21:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 05 03:20:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	357113	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	349830	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	188471	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	96492	45.199	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.400%
7) Chloroethane-d5	1.532	69	68356	45.955	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.900%
11) 1,1-Dichloroethene-d2	2.060	65	36815	45.331	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.660%
21) 2-Butanone-d5	3.777	46	104244	98.313	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.310%
24) Chloroform-d	4.246	84	201912	44.567	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.140%
26) 1,2-Dichloroethane-d4	4.937	65	132222	46.941	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.880%
32) Benzene-d6	4.957	84	388640	47.501	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.000%
36) 1,2-Dichloropropane-d6	5.985	67	107687	48.951	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.900%
41) Toluene-d8	7.243	98	362740	46.476	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.960%
43) trans-1,3-Dichloroprop...	7.551	79	62488	44.921	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.840%
47) 2-Hexanone-d5	8.021	63	83775	94.793	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.790%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	145234	46.442	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.880%
66) 1,2-Dichlorobenzene-d4	11.561	152	166888	49.768	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.540%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120423\
Data File : VV033089.D
Acq On : 04 Dec 2023 10:08
Operator : SY/MD
Sample : VV1204MBL01
Misc : 5.0G/5.0mL/100uL/5mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK230

Quant Time: Dec 05 03:21:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 05 03:20:48 2023
Response via : Initial Calibration

