

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120425\
 Data File : VV039469.D
 Acq On : 04 Dec 2025 15:24
 Operator : SY/MD
 Sample : VIBLK229
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK229

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/10/2025
 Supervised By : Mahesh Dadoda 12/10/2025

Quant Time: Dec 05 00:10:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Dec 04 00:58:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	6254	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	6214	0.500	ug/L	0.02
12) 1,4-Dichlorobenzene-d4	11.197	152	3212	0.500	ug/L	0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.294	65	3477m	0.553	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	110.000%		
4) 1,2-Dichloroethane-d4	4.993	65	1933	0.464	ug/L	0.02
Spiked Amount 0.500	Range 70 - 130		Recovery =	92.000%		
7) 1,2-Dichloropropane-d6	5.995	67	2215	0.422	ug/L	-0.02
Spiked Amount 0.500	Range 60 - 140		Recovery =	84.000%		
8) Toluene-d8	7.253	98	4155	0.470	ug/L	0.02
Spiked Amount 0.500	Range 70 - 130		Recovery =	94.000%		
10) 1,1,2,2-Tetrachloroeth...	10.133	84	1606	0.544	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	108.000%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120425\
Data File : VV039469.D
Acq On : 04 Dec 2025 15:24
Operator : SY/MD
Sample : VIBLK229
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK229

Quant Time: Dec 05 00:10:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Dec 04 00:58:46 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 12/10/2025
Supervised By : Mahesh Dadoda 12/10/2025

