

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120425\
 Data File : VV039471.D
 Acq On : 04 Dec 2025 16:56
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTDCCC0.5EC

Quant Time: Dec 05 00:26:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Dec 04 00:58:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	6921	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.785	117	7314	0.500	ug/L	0.00
12) 1,4-Dichlorobenzene-d4	11.175	152	4168	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.295	65	3966	0.570	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	114.000%
4) 1,2-Dichloroethane-d4	4.970	65	2331	0.506	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.000%
7) 1,2-Dichloropropane-d6	6.020	67	3374	0.547	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	110.000%
8) Toluene-d8	7.254	98	5330	0.512	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.000%
10) 1,1,2,2-Tetrachloroeth...	10.133	84	1967	0.566	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	114.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.295	62	7682	0.513	ug/L	# 1
6) Trichloroethene	5.853	95	4720	0.521	ug/L	99
9) 1,2-Dibromoethane	8.290	107	3092	0.516	ug/L	100
11) 1,1,2,2-Tetrachloroethane	10.155	83	4141	0.564	ug/L	99
13) 1,2,3-Trichloropropane	10.200	75	2302	0.525	ug/L	97
14) 1,2-Dibromo-3-chloropr...	12.351	75	447	0.503	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120425\
Data File : VV039471.D
Acq On : 04 Dec 2025 16:56
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTDCCC0.5EC

Quant Time: Dec 05 00:26:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Dec 04 00:58:46 2025
Response via : Initial Calibration

