

Data File : VV013927.D
Acq On : 05 Dec 2019 20:03
Operator : SY/MD
Sample : K6170-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Dec 06 05:53:54 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 06 04:06:40 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	597035	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	574182	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	221478	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	200278	58.00	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	116.00%		
7) Chloroethane-d5	1.57	69	202027	56.05	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	112.10%		
11) 1,1-Dichloroethene-d2	2.13	63	304890m	39.93	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	79.86%		
21) 2-Butanone-d5	3.92	46	243655	100.25	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	100.25%		
24) Chloroform-d	4.40	84	362203	49.95	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	99.90%		
26) 1,2-Dichloroethane-d4	5.08	65	231625	50.41	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	100.82%		
32) Benzene-d6	5.09	84	733553	49.30	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	98.60%		
36) 1,2-Dichloropropane-d6	6.11	67	226593	50.74	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	101.48%		
41) Toluene-d8	7.35	98	670437	49.43	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	98.86%		
43) trans-1,3-Dichloropropene-	7.66	79	101565	46.74	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	93.48%		
47) 2-Hexanone-d5	8.13	63	147724	88.58	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	88.58%		
57) 1,1,2,2-Tetrachloroethane-	10.25	84	283556	43.85	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	87.70%		
64) 1,2-Dichlorobenzene-d4	11.67	152	236955	54.47	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	108.94%		

Target Compounds					Qvalue
13) Acetone	2.18	43	5158	1.448 ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV013927.D
Acq On : 05 Dec 2019 20:03
Operator : SY/MD
Sample : K6170-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Dec 06 05:53:54 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 06 04:06:40 2019
Response via : Initial Calibration

