

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV120519\
 Data File : VV013912.D
 Acq On : 05 Dec 2019 13:34
 Operator : SY/MD
 Sample : K6165-05
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBBT2

Quant Time: Dec 06 04:24:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 06 04:06:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	782512	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	690354	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	274098	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	166998	36.90	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	73.80%
7) Chloroethane-d5	1.58	69	195795	41.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.90%
11) 1,1-Dichloroethene-d2	2.13	63	290042	28.98	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	57.96%#
21) 2-Butanone-d5	3.92	46	280232	87.97	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	87.97%
24) Chloroform-d	4.39	84	410785	43.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.44%
26) 1,2-Dichloroethane-d4	5.07	65	262396	43.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.14%
32) Benzene-d6	5.09	84	804171	44.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.90%
36) 1,2-Dichloropropane-d6	6.11	67	251145	46.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.54%
41) Toluene-d8	7.35	98	679875	41.69	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.38%
43) trans-1,3-Dichloropropene-	7.66	79	109356	41.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.70%
47) 2-Hexanone-d5	8.13	63	148748	74.19	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	74.19%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	318279	40.93	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.86%
64) 1,2-Dichlorobenzene-d4	11.67	152	261890	48.65	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.30%

Target Compounds					Ovalue
13) Acetone	2.18	43	5839	1.251 ug/L	88
70) 1,2,3-Trichlorobenzene	13.79	180	6245	0.906 ug/L #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV120519\
Data File : VV013912.D
Acq On : 05 Dec 2019 13:34
Operator : SY/MD
Sample : K6165-05
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBBT2

Quant Time: Dec 06 04:24:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 06 04:06:40 2019
Response via : Initial Calibration

