

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV120519\
 Data File : VV013918.D
 Acq On : 05 Dec 2019 16:00
 Operator : SY/MD
 Sample : K6165-08
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBBW3

Quant Time: Dec 06 04:46:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 06 04:06:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	576154	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	581184	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	230556	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	135459	40.65	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.30%
7) Chloroethane-d5	1.57	69	157592	45.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.62%
11) 1,1-Dichloroethene-d2	2.13	63	234680	31.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.70%
21) 2-Butanone-d5	3.92	46	263016	112.14	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	112.14%
24) Chloroform-d	4.39	84	343218	49.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.10%
26) 1,2-Dichloroethane-d4	5.07	65	222406	50.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.32%
32) Benzene-d6	5.09	84	671995	44.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.24%
36) 1,2-Dichloropropane-d6	6.11	67	217677	48.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.30%
41) Toluene-d8	7.35	98	568180	41.38	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.76%
43) trans-1,3-Dichloropropene-	7.66	79	94890	43.14	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.28%
47) 2-Hexanone-d5	8.13	63	152314	90.23	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.23%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	323791	49.47	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.94%
64) 1,2-Dichlorobenzene-d4	11.67	152	247086	54.57	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.14%

Target Compounds

					Ovalue
13) Acetone	2.18	43	13118	3.817 ug/L	93
25) Chloroform	4.42	83	115346	14.628 ug/L	99
38) Bromodichloromethane	6.55	83	51010	8.587 ug/L	99
40) 4-Methyl-2-pentanone	7.28	43	5817	1.068 ug/L #	55
46) Tetrachloroethene	8.02	164	3192	0.762 ug/L	88
49) Dibromochloromethane	8.29	129	16703	3.286 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV120519\
Data File : VV013918.D
Acq On : 05 Dec 2019 16:00
Operator : SY/MD
Sample : K6165-08
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBBW3

Quant Time: Dec 06 04:46:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 06 04:06:40 2019
Response via : Initial Calibration

