

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120525\
 Data File : VV039477.D
 Acq On : 05 Dec 2025 14:16
 Operator : SY/MD
 Sample : Q3757-08
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 RMW-05B-90.5-120325-SIM

Quant Time: Dec 05 22:48:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Dec 04 00:58:46 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	6011	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	5568	0.500	ug/L	0.02
12) 1,4-Dichlorobenzene-d4	11.197	152	2844	0.500	ug/L	0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.294	65	3092	0.512	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	102.000%
4) 1,2-Dichloroethane-d4	4.969	65	1638	0.409	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	82.000%
7) 1,2-Dichloropropane-d6	5.995	67	1778	0.378	ug/L	-0.02
Spiked Amount	0.500	Range	60 - 140	Recovery	=	76.000%
8) Toluene-d8	7.253	98	3360	0.424	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	84.000%
10) 1,1,2,2-Tetrachloroeth...	10.133	84	1359	0.513	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	102.000%
Target Compounds						
						Qvalue
3) Vinyl chloride	1.294	62	1523	0.117	ug/L	# 1
6) Trichloroethene	5.852	95	2319	0.336	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120525\
Data File : VV039477.D
Acq On : 05 Dec 2025 14:16
Operator : SY/MD
Sample : Q3757-08
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
RMW-05B-90.5-120325-SIM

Quant Time: Dec 05 22:48:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Dec 04 00:58:46 2025
Response via : Initial Calibration

