

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120525\
 Data File : VV039479.D
 Acq On : 05 Dec 2025 14:59
 Operator : SY/MD
 Sample : Q3757-04 20X
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BR-03D-490-120325-SIM

Quant Time: Dec 05 22:49:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Dec 04 00:58:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	5792	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	5769	0.500	ug/L	0.02
12) 1,4-Dichlorobenzene-d4	11.197	152	2978	0.500	ug/L	0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.294	65	3452	0.593	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	118.000%
4) 1,2-Dichloroethane-d4	4.993	65	1884	0.488	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	98.000%
7) 1,2-Dichloropropane-d6	5.996	67	2047	0.421	ug/L	-0.02
Spiked Amount	0.500	Range	60 - 140	Recovery	=	84.000%
8) Toluene-d8	7.254	98	3945	0.480	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	96.000%
10) 1,1,2,2-Tetrachloroeth...	10.133	84	1506	0.549	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	110.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.294	62	2923	0.233	ug/L #	1
6) Trichloroethene	5.852	95	10862	1.519	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120525\
Data File : VV039479.D
Acq On : 05 Dec 2025 14:59
Operator : SY/MD
Sample : Q3757-04 20X
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BR-03D-490-120325-SIM

Quant Time: Dec 05 22:49:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Dec 04 00:58:46 2025
Response via : Initial Calibration

