

Data File : VV013937.D
Acq On : 06 Dec 2019 12:59
Operator : SY/MD
Sample : K6170-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Dec 07 02:37:53 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 07 01:39:07 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	490337	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	473191	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	178685	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	191501	67.53	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	135.06%#		
7) Chloroethane-d5	1.57	69	195950	66.20	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	132.40%#		
11) 1,1-Dichloroethene-d2	2.13	63	304648m	48.58	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	97.16%		
21) 2-Butanone-d5	3.92	46	235490	117.98	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	117.98%		
24) Chloroform-d	4.39	84	352212	59.14	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	118.28%		
26) 1,2-Dichloroethane-d4	5.07	65	226122	59.92	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	119.84%		
32) Benzene-d6	5.09	84	734066	59.87	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	119.74%		
36) 1,2-Dichloropropane-d6	6.11	67	221360	60.14	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	120.28%#		
41) Toluene-d8	7.35	98	624343	55.85	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	111.70%		
43) trans-1,3-Dichloropropene-	7.66	79	95257	53.19	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	106.38%		
47) 2-Hexanone-d5	8.13	63	134662	97.98	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	97.98%		
57) 1,1,2,2-Tetrachloroethane-	10.26	84	267087	50.12	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	100.24%		
64) 1,2-Dichlorobenzene-d4	11.67	152	223894	63.80	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	127.60%#		

Target Compounds					Qvalue
13) Acetone	2.19	43	3000	1.026 ug/L	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV013937.D
Acq On : 06 Dec 2019 12:59
Operator : SY/MD
Sample : K6170-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Dec 07 02:37:53 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 07 01:39:07 2019
Response via : Initial Calibration

