

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV120619\
 Data File : VV013943.D
 Acq On : 06 Dec 2019 15:26
 Operator : SY/MD
 Sample : K6170-12
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BT8

Quant Time: Dec 07 03:04:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 07 01:39:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	598585	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	584737	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	220453	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	189927	54.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	109.72%
7) Chloroethane-d5	1.56	69	194620	53.86	ug/L	-0.01
Spiked Amount	50.000	Range	70 - 130	Recovery	=	107.72%
11) 1,1-Dichloroethene-d2	2.12	63	301531	39.39	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.78%
21) 2-Butanone-d5	3.92	46	225169	92.41	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.41%
24) Chloroform-d	4.39	84	351219	48.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.62%
26) 1,2-Dichloroethane-d4	5.07	65	223139	48.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.88%
32) Benzene-d6	5.09	84	713029	47.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.12%
36) 1,2-Dichloropropane-d6	6.11	67	220874	48.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.12%
41) Toluene-d8	7.35	98	622319	45.05	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.10%
43) trans-1,3-Dichloropropene-	7.66	79	92219	41.67	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.34%
47) 2-Hexanone-d5	8.13	63	124569	73.35	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	73.35%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	276595	42.00	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	224136	51.77	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.54%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV120619\
Data File : VV013943.D
Acq On : 06 Dec 2019 15:26
Operator : SY/MD
Sample : K6170-12
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BT8

Quant Time: Dec 07 03:04:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 07 01:39:07 2019
Response via : Initial Calibration

