

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120622\
 Data File : VV029405.D
 Acq On : 06 Dec 2022 11:04
 Operator : SY/MD
 Sample : VV1206WBL07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK286

Quant Time: Dec 06 21:49:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Dec 06 21:48:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	216695	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	201122	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	103652	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	65856	4.938	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.800%
7) Chloroethane-d5	1.564	69	50462	5.169	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.400%
11) 1,1-Dichloroethene-d2	2.105	63	76951	3.606	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.200%
20) 2-Butanone-d5	3.899	46	81430	36.502	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	73.000%
24) Chloroform-d	4.339	84	111441	4.682	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.600%
26) 1,2-Dichloroethane-d4	5.027	65	48994	4.678	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
32) Benzene-d6	5.047	84	237948	4.281	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.600%
36) 1,2-Dichloropropane-d6	6.063	67	70357	4.504	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.000%
41) Toluene-d8	7.310	98	215918	4.116	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.400%
43) trans-1,3-Dichloroprop...	7.619	79	25641	4.258	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.200%
46) 2-Hexanone-d5	8.085	63	91055	42.419	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.840%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	47129	4.890	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.800%
66) 1,2-Dichlorobenzene-d4	11.612	152	78332	4.326	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120622\
Data File : VV029405.D
Acq On : 06 Dec 2022 11:04
Operator : SY/MD
Sample : VV1206WBL07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK286

Quant Time: Dec 06 21:49:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Dec 06 21:48:39 2022
Response via : Initial Calibration

