

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV120718\
 Data File : VV008876.D
 Acq On : 07 Dec 2018 15:09
 Operator : SY/MD
 Sample : VV1207WBL01
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK62

Quant Time: Dec 08 00:30:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR120518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 07 22:19:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	114717	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	104985	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	45602	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	27850	4.55	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.00%
7) Chloroethane-d5	1.58	69	22081	4.89	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.80%
11) 1,1-Dichloroethene-d2	2.13	63	49924	4.23	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.60%
20) 2-Butanone-d5	3.98	46	78028	51.64	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.28%
24) Chloroform-d	4.40	84	68575	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
26) 1,2-Dichloroethane-d4	5.08	65	35104	5.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.40%
32) Benzene-d6	5.10	84	140827	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.60%
36) 1,2-Dichloropropane-d6	6.12	67	41372	5.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.60%
41) Toluene-d8	7.36	98	124392	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.80%
43) trans-1,3-Dichloropropene-	7.67	79	15636	5.14	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	102.80%
46) 2-Hexanone-d5	8.14	63	60874	52.52	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.04%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	26694	4.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	45697	5.44	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV120718\
Data File : VV008876.D
Acq On : 07 Dec 2018 15:09
Operator : SY/MD
Sample : VV1207WBL01
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK62

Quant Time: Dec 08 00:30:32 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR120518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 07 22:19:33 2018
Response via : Initial Calibration

