

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV120720\
 Data File : VV019494.D
 Acq On : 07 Dec 2020 16:20
 Operator : SY/MD
 Sample : VV1207MBL01
 Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK96

Quant Time: Dec 08 03:18:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120720WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 08 03:05:02 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	328068	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	320130	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	139310	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	106963	52.37	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	104.74%
7) Chloroethane-d5	1.58	69	102344	49.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.78%
11) 1,1-Dichloroethene-d2	2.12	63	194051	41.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.12%
21) 2-Butanone-d5	3.90	46	152206	97.48	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.48%
24) Chloroform-d	4.37	84	239654	50.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.30%
26) 1,2-Dichloroethane-d4	5.05	65	170054	53.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.98%
32) Benzene-d6	5.07	84	472251	52.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.58%
36) 1,2-Dichloropropane-d6	6.09	67	161223	52.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.86%
41) Toluene-d8	7.33	98	395439	51.22	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.44%
43) trans-1,3-Dichloropropene-	7.64	79	71450	49.15	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.30%
47) 2-Hexanone-d5	8.11	63	79931	89.47	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.47%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	166282	46.00	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.00%
64) 1,2-Dichlorobenzene-d4	11.64	152	140089	55.43	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.86%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV120720\
Data File : VV019494.D
Acq On : 07 Dec 2020 16:20
Operator : SY/MD
Sample : VV1207MBL01
Misc : 5.0µ/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK96

Quant Time: Dec 08 03:18:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120720WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 08 03:05:02 2020
Response via : Initial Calibration

