

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV120720\
 Data File : VV019503.D
 Acq On : 07 Dec 2020 20:35
 Operator : SY/MD
 Sample : VIBLK78
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK78

Quant Time: Dec 08 04:01:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120720WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 08 03:05:02 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	289966	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	283562	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	120398	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	67345	37.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	74.62%
7) Chloroethane-d5	1.58	69	91274	49.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.68%
11) 1,1-Dichloroethene-d2	2.12	63	182789	43.76	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.52%
21) 2-Butanone-d5	3.90	46	149844	108.58	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.58%
24) Chloroform-d	4.37	84	227675	54.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.88%
26) 1,2-Dichloroethane-d4	5.05	65	162749	57.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	115.84%
32) Benzene-d6	5.07	84	447584	56.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.96%
36) 1,2-Dichloropropane-d6	6.09	67	151523	56.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	112.32%
41) Toluene-d8	7.33	98	371690	54.35	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.70%
43) trans-1,3-Dichloropropene-	7.64	79	66356	51.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	103.06%
47) 2-Hexanone-d5	8.10	63	76727	96.96	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.96%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	158389	49.46	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.92%
64) 1,2-Dichlorobenzene-d4	11.65	152	131917	60.40	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	120.80%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV120720\
Data File : VV019503.D
Acq On : 07 Dec 2020 20:35
Operator : SY/MD
Sample : VIBLK78
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK78

Quant Time: Dec 08 04:01:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120720WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 08 03:05:02 2020
Response via : Initial Calibration

