

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV120721\  
 Data File : VV023811.D  
 Acq On : 07 Dec 2021 10:25  
 Operator : SY/MD  
 Sample : VV1207WBL01  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VBLK271

Quant Time: Dec 08 01:14:19 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR112321WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Dec 02 02:08:23 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon           | Response | Conc   | Units    | Dev(Min) |
|-------------------------------|--------|----------------|----------|--------|----------|----------|
| -----                         |        |                |          |        |          |          |
| Internal Standards            |        |                |          |        |          |          |
| 1) 1,4-Difluorobenzene        | 5.613  | 114            | 154123   | 5.000  | ug/L     | 0.00     |
| 28) Chlorobenzene-d5          | 8.850  | 117            | 147675   | 5.000  | ug/L     | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.249 | 152            | 74057    | 5.000  | ug/L     | 0.00     |
| System Monitoring Compounds   |        |                |          |        |          |          |
| 4) Vinyl Chloride-d3          | 1.304  | 65             | 59954    | 4.739  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 40 - 130 | Recovery | =      | 94.800%  |          |
| 7) Chloroethane-d5            | 1.568  | 69             | 44541    | 4.479  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 130 | Recovery | =      | 89.600%  |          |
| 11) 1,1-Dichloroethene-d2     | 2.105  | 63             | 80111    | 3.593  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 125 | Recovery | =      | 71.800%  |          |
| 20) 2-Butanone-d5             | 3.899  | 46             | 63921    | 42.026 | ug/L     | 0.00     |
| Spiked Amount                 | 50.000 | Range 40 - 130 | Recovery | =      | 84.060%  |          |
| 24) Chloroform-d              | 4.342  | 84             | 104486   | 4.743  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 94.800%  |          |
| 26) 1,2-Dichloroethane-d4     | 5.027  | 65             | 47363    | 4.602  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 92.000%  |          |
| 32) Benzene-d6                | 5.043  | 84             | 207362   | 5.155  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 103.000% |          |
| 36) 1,2-Dichloropropane-d6    | 6.066  | 67             | 58415    | 5.180  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 140 | Recovery | =      | 103.600% |          |
| 41) Toluene-d8                | 7.313  | 98             | 188792   | 5.023  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 100.400% |          |
| 43) trans-1,3-Dichloroprop... | 7.622  | 79             | 22306    | 4.907  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 55 - 130 | Recovery | =      | 98.200%  |          |
| 46) 2-Hexanone-d5             | 8.088  | 63             | 87311    | 57.808 | ug/L     | 0.00     |
| Spiked Amount                 | 50.000 | Range 45 - 130 | Recovery | =      | 115.620% |          |
| 56) 1,1,2,2-Tetrachloroeth... | 10.217 | 84             | 38853    | 4.788  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 120 | Recovery | =      | 95.800%  |          |
| 66) 1,2-Dichlorobenzene-d4    | 11.622 | 152            | 71362    | 5.450  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 80 - 120 | Recovery | =      | 109.000% |          |
| Target Compounds              |        |                |          |        |          |          |
| 42) Toluene                   | 7.397  | 91             | 4345     | 0.095  | ug/L     | 98       |
| -----                         |        |                |          |        |          |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV120721\  
Data File : VV023811.D  
Acq On : 07 Dec 2021 10:25  
Operator : SY/MD  
Sample : VV1207WBL01  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VBLK271

Quant Time: Dec 08 01:14:19 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR112321WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Thu Dec 02 02:08:23 2021  
Response via : Initial Calibration

