

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120721\
 Data File : VW023814.D
 Acq On : 07 Dec 2021 11:45
 Operator : SY/MD
 Sample : VIBLK243
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK243

Quant Time: Dec 08 01:14:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 02 02:08:23 2021
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.616	114	144765	5.000	ug/L	0.00
28) Chlorobenzene-d5		8.850	117	138543	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.249	152	68367	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.307	65	60454	5.087	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.800%	
7) Chloroethane-d5		1.568	69	45345	4.854	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.000%	
11) 1,1-Dichloroethene-d2		2.108	63	79697	3.805	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.200%	
20) 2-Butanone-d5		3.908	46	60033	42.021	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.040%	
24) Chloroform-d		4.346	84	103547	5.004	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.000%	
26) 1,2-Dichloroethane-d4		5.030	65	48391	5.005	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.200%	
32) Benzene-d6		5.047	84	207451	5.497	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.000%	
36) 1,2-Dichloropropane-d6		6.066	67	58445	5.524	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.400%	
41) Toluene-d8		7.313	98	189483	5.374	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.400%	
43) trans-1,3-Dichloroprop...		7.622	79	23056	5.406	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	108.200%	
46) 2-Hexanone-d5		8.088	63	91904	64.860	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	129.720%	
56) 1,1,2,2-Tetrachloroeth...		10.214	84	39229	5.152	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.000%	
66) 1,2-Dichlorobenzene-d4		11.625	152	71082	5.881	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.600%	
Target Compounds							
16) Methylene chloride		2.506	84	8149	0.589	ug/L	87
42) Toluene		7.397	91	2930	0.068	ug/L	83
47) Tetrachloroethene		7.979	164	795	0.083	ug/L	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120721\
Data File : VV023814.D
Acq On : 07 Dec 2021 11:45
Operator : SY/MD
Sample : VIBLK243
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK243

Quant Time: Dec 08 01:14:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 02 02:08:23 2021
Response via : Initial Calibration

