

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120721\
 Data File : VW023821.D
 Acq On : 07 Dec 2021 14:31
 Operator : SY/MD
 Sample : M4879-12DL 50X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0G69DL

Quant Time: Dec 08 01:15:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 02 02:08:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	129534	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	124059	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	60602	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	56456	5.309	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.200%
7) Chloroethane-d5	1.568	69	41969	5.021	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.400%
11) 1,1-Dichloroethene-d2	2.108	63	75306	4.018	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.400%
20) 2-Butanone-d5	3.908	46	52330	40.936	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.880%
24) Chloroform-d	4.349	84	98662	5.328	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.600%
26) 1,2-Dichloroethane-d4	5.034	65	47937	5.542	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.800%
32) Benzene-d6	5.050	84	193738	5.733	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	114.600%
36) 1,2-Dichloropropane-d6	6.069	67	54187	5.719	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	114.400%
41) Toluene-d8	7.317	98	178561	5.655	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.000%
43) trans-1,3-Dichloroprop...	7.622	79	21438	5.614	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	112.200%
46) 2-Hexanone-d5	8.091	63	86727	68.352	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	136.700%#
56) 1,1,2,2-Tetrachloroeth...	10.217	84	36940	5.418	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.400%
66) 1,2-Dichlorobenzene-d4	11.625	152	67668	6.316	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	126.400%#
Target Compounds						Qvalue
22) cis-1,2-Dichloroethene	3.915	96	1965	0.207	ug/L	84
34) Trichloroethene	5.921	95	1841	0.194	ug/L	81
42) Toluene	7.400	91	3076	0.080	ug/L	96
47) Tetrachloroethene	7.976	164	101660	11.789	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120721\
Data File : VV023821.D
Acq On : 07 Dec 2021 14:31
Operator : SY/MD
Sample : M4879-12DL 50X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0G69DL

Quant Time: Dec 08 01:15:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 02 02:08:23 2021
Response via : Initial Calibration

