

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120722\
 Data File : VW029450.D
 Acq On : 08 Dec 2022 04:33
 Operator : SY/MD
 Sample : N5919-13
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0EG7

Quant Time: Dec 08 05:39:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 08 01:48:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	477619	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	448237	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	231238	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	121859	41.665	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.320%
7) Chloroethane-d5	1.564	69	101732	45.616	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.240%
11) 1,1-Dichloroethene-d2	2.105	63	155236	34.030	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.060%
21) 2-Butanone-d5	3.870	46	184559	103.443	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.440%
24) Chloroform-d	4.339	84	255009	46.333	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.660%
26) 1,2-Dichloroethane-d4	5.024	65	154791	49.243	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.480%
32) Benzene-d6	5.043	84	549880	46.427	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.860%
36) 1,2-Dichloropropane-d6	6.063	67	171490	47.402	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.800%
41) Toluene-d8	7.310	98	501786	44.964	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.920%
43) trans-1,3-Dichloroprop...	7.612	79	79013	43.251	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.500%
47) 2-Hexanone-d5	8.082	63	158040	102.300	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.300%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	247336	47.843	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.680%
66) 1,2-Dichlorobenzene-d4	11.612	152	211010	47.936	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120722\
Data File : VV029450.D
Acq On : 08 Dec 2022 04:33
Operator : SY/MD
Sample : N5919-13
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0EG7

Quant Time: Dec 08 05:39:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 08 01:48:46 2022
Response via : Initial Calibration

