

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120722\
 Data File : VW029455.D
 Acq On : 08 Dec 2022 06:31
 Operator : SY/MD
 Sample : N5919-18
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0EH4

Quant Time: Dec 08 07:51:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 08 01:48:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	474340	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	436578	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	223596	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	120980	41.650	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.300%
7) Chloroethane-d5	1.564	69	101360	45.763	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.520%
11) 1,1-Dichloroethene-d2	2.104	63	153205	33.816	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.640%
21) 2-Butanone-d5	3.870	46	189625	107.017	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.020%
24) Chloroform-d	4.339	84	254579	46.575	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.140%
26) 1,2-Dichloroethane-d4	5.024	65	154871	49.609	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.220%
32) Benzene-d6	5.043	84	542590	47.035	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.060%
36) 1,2-Dichloropropane-d6	6.063	67	170449	48.373	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.740%
41) Toluene-d8	7.310	98	498397	45.853	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.700%
43) trans-1,3-Dichloroprop...	7.612	79	79250	44.539	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.080%
47) 2-Hexanone-d5	8.082	63	157643	104.768	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	104.770%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	245782	48.812	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.620%
66) 1,2-Dichlorobenzene-d4	11.612	152	207132	48.664	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120722\
Data File : VV029455.D
Acq On : 08 Dec 2022 06:31
Operator : SY/MD
Sample : N5919-18
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0EH4

Quant Time: Dec 08 07:51:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 08 01:48:46 2022
Response via : Initial Calibration

