

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120722\
 Data File : VW029457.D
 Acq On : 08 Dec 2022 07:18
 Operator : SY/MD
 Sample : N5920-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0EH9

Quant Time: Dec 08 07:56:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 08 01:48:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	493793	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	458278	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	235687	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	122341	40.460	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	80.920%		
7) Chloroethane-d5	1.564	69	100646	43.651	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	87.300%		
11) 1,1-Dichloroethene-d2	2.104	63	153526	32.552	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	65.100%		
21) 2-Butanone-d5	3.866	46	184243	99.883	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	99.880%		
24) Chloroform-d	4.339	84	256911	45.150	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	90.300%		
26) 1,2-Dichloroethane-d4	5.024	65	155345	47.800	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	95.600%		
32) Benzene-d6	5.043	84	544723	44.984	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	89.960%		
36) 1,2-Dichloropropane-d6	6.063	67	169846	45.919	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	91.840%		
41) Toluene-d8	7.307	98	498038	43.651	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	87.300%		
43) trans-1,3-Dichloroprop...	7.616	79	78334	41.940	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	83.880%		
47) 2-Hexanone-d5	8.082	63	153540	97.209	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	97.210%		
56) 1,1,2,2-Tetrachloroeth...	10.207	84	243990	46.161	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	92.320%		
66) 1,2-Dichlorobenzene-d4	11.612	152	209009	46.585	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	93.180%		
Target Compounds						Qvalue
34) Trichloroethene	5.915	95	5814	1.796	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120722\
Data File : VV029457.D
Acq On : 08 Dec 2022 07:18
Operator : SY/MD
Sample : N5920-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0EH9

Quant Time: Dec 08 07:56:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 08 01:48:46 2022
Response via : Initial Calibration

