

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120722\
 Data File : VW029428.D
 Acq On : 07 Dec 2022 19:31
 Operator : SY/MD
 Sample : N5894-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AA0

Quant Time: Dec 08 00:57:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 08 00:46:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	518797	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	468371	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	239209	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	134872	42.454	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.900%
7) Chloroethane-d5	1.561	69	105385	43.503	ug/L	0.01
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.000%
11) 1,1-Dichloroethene-d2	2.105	63	161537	32.600	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.200%
21) 2-Butanone-d5	3.870	46	190252	98.170	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.170%
24) Chloroform-d	4.339	84	262726	43.946	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.900%
26) 1,2-Dichloroethane-d4	5.024	65	160919	47.129	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.260%
32) Benzene-d6	5.043	84	569160	45.989	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.980%
36) 1,2-Dichloropropane-d6	6.063	67	175888	46.528	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.060%
41) Toluene-d8	7.307	98	519142	44.520	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.040%
43) trans-1,3-Dichloroprop...	7.613	79	84347	44.186	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.380%
47) 2-Hexanone-d5	8.082	63	164700	102.028	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.030%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	255110	47.225	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.460%
66) 1,2-Dichlorobenzene-d4	11.612	152	212140	46.587	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.180%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120722\
Data File : VV029428.D
Acq On : 07 Dec 2022 19:31
Operator : SY/MD
Sample : N5894-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AA0

Quant Time: Dec 08 00:57:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 08 00:46:44 2022
Response via : Initial Calibration

