

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120723\
 Data File : VW033145.D
 Acq On : 07 Dec 2023 14:39
 Operator : SY/MD
 Sample : 05673-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHEJO

Quant Time: Dec 08 01:38:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 08 01:34:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	296753	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	294020	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	151773	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	57723	32.539	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	65.080%
7) Chloroethane-d5	1.532	69	46403	37.541	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	75.080%
11) 1,1-Dichloroethene-d2	2.060	65	23854	35.346	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.700%
21) 2-Butanone-d5	3.783	46	76464	86.782	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.780%
24) Chloroform-d	4.249	84	153791	40.850	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.700%
26) 1,2-Dichloroethane-d4	4.941	65	102072	43.608	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.220%
32) Benzene-d6	4.960	84	273287	39.742	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.480%
36) 1,2-Dichloropropane-d6	5.989	67	77765	42.060	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	84.120%
41) Toluene-d8	7.243	98	251557	38.349	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	76.700%#
43) trans-1,3-Dichloroprop...	7.555	79	44275	37.869	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.740%
47) 2-Hexanone-d5	8.024	63	66352	89.330	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.330%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	111398	42.384	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.760%
66) 1,2-Dichlorobenzene-d4	11.561	152	125010	46.293	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.580%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120723\
Data File : VV033145.D
Acq On : 07 Dec 2023 14:39
Operator : SY/MD
Sample : 05673-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHEJ0

Quant Time: Dec 08 01:38:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 08 01:34:57 2023
Response via : Initial Calibration

