

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120723\
 Data File : VW033150.D
 Acq On : 07 Dec 2023 16:35
 Operator : SY/MD
 Sample : 05645-13
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0G67

Quant Time: Dec 08 01:39:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 08 01:34:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	287042	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	282975	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	147222	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	54505	31.764	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	63.520%
7) Chloroethane-d5	1.532	69	44195	36.965	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	73.920%
11) 1,1-Dichloroethene-d2	2.060	65	23478	35.966	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.940%
21) 2-Butanone-d5	3.780	46	82581	96.895	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.900%
24) Chloroform-d	4.249	84	150173	41.239	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.480%
26) 1,2-Dichloroethane-d4	4.941	65	97399	43.020	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.040%
32) Benzene-d6	4.960	84	265150	40.064	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.120%
36) 1,2-Dichloropropane-d6	5.989	67	74465	41.847	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.700%
41) Toluene-d8	7.243	98	241810	38.301	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	76.600%#
43) trans-1,3-Dichloroprop...	7.551	79	42340	37.628	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.260%
47) 2-Hexanone-d5	8.021	63	64028	89.566	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.570%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	115000	45.462	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.920%
66) 1,2-Dichlorobenzene-d4	11.561	152	120653	46.061	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.120%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120723\
Data File : VV033150.D
Acq On : 07 Dec 2023 16:35
Operator : SY/MD
Sample : 05645-13
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0G67

Quant Time: Dec 08 01:39:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 08 01:34:57 2023
Response via : Initial Calibration

