

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120723\
 Data File : VW033160.D
 Acq On : 07 Dec 2023 20:50
 Operator : SY/MD
 Sample : 05689-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0G76

Quant Time: Dec 08 04:16:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 08 04:15:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	298407	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	295232	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	157529	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	54276	30.426	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	60.860%
7) Chloroethane-d5	1.532	69	45146	36.322	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	72.640%
11) 1,1-Dichloroethene-d2	2.060	65	23371	34.439	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.880%
21) 2-Butanone-d5	3.780	46	84988	95.921	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.920%
24) Chloroform-d	4.246	84	151807	40.100	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.200%
26) 1,2-Dichloroethane-d4	4.940	65	101030	42.924	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.840%
32) Benzene-d6	4.960	84	271334	39.296	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.600%
36) 1,2-Dichloropropane-d6	5.989	67	77266	41.618	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.240%
41) Toluene-d8	7.243	98	242955	36.885	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	73.780%#
43) trans-1,3-Dichloroprop...	7.555	79	43907	37.400	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.800%
47) 2-Hexanone-d5	8.024	63	67734	90.816	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.820%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	115822	43.886	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.780%
66) 1,2-Dichlorobenzene-d4	11.561	152	124219	44.319	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120723\
Data File : VV033160.D
Acq On : 07 Dec 2023 20:50
Operator : SY/MD
Sample : 05689-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0G76

Quant Time: Dec 08 04:16:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 08 04:15:58 2023
Response via : Initial Calibration

