

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120821\
 Data File : VW023840.D
 Acq On : 08 Dec 2021 16:33
 Operator : SY/MD
 Sample : M4889-16
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0CN7

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/10/2021
 Supervised By : Mahesh Dadoda 12/10/2021

Quant Time: Dec 09 00:34:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 02 02:08:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	108193	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	101619	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	51470	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	38625	4.349	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	87.000%	
7) Chloroethane-d5	1.564	69	31233	4.474	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	89.400%	
11) 1,1-Dichloroethene-d2	2.105	63	52150	3.332	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	66.600%	
20) 2-Butanone-d5	3.912	46	56136m	52.575	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	105.160%	
24) Chloroform-d	4.349	84	76718	4.960	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.200%	
26) 1,2-Dichloroethane-d4	5.034	65	36971	5.117	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	102.400%	
32) Benzene-d6	5.050	84	140637	5.081	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.600%	
36) 1,2-Dichloropropane-d6	6.069	67	41760	5.381	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	107.600%	
41) Toluene-d8	7.317	98	125608	4.856	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.200%	
43) trans-1,3-Dichloroprop...	7.625	79	16813	5.375	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	107.400%	
46) 2-Hexanone-d5	8.092	63	72703	69.953	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	139.900%#	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	30489	5.460	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	109.200%	
66) 1,2-Dichlorobenzene-d4	11.625	152	51553	5.665	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	113.400%	
Target Compounds					Qvalue	
42) Toluene	7.400	91	2389	0.076	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120821\
Data File : VV023840.D
Acq On : 08 Dec 2021 16:33
Operator : SY/MD
Sample : M4889-16
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0CN7

Quant Time: Dec 09 00:34:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 02 02:08:23 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 12/10/2021
Supervised By : Mahesh Dadoda 12/10/2021

