

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120822\
 Data File : VW029487.D
 Acq On : 08 Dec 2022 20:56
 Operator : SY/MD
 Sample : N5920-09
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0EJ8

Quant Time: Dec 09 04:15:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 08 01:48:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	520964	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	478385	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	250215	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	109417	34.298	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	68.600%
7) Chloroethane-d5	1.564	69	101200	41.602	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.200%
11) 1,1-Dichloroethene-d2	2.105	63	154140	30.978	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.960%
21) 2-Butanone-d5	3.867	46	205045	105.363	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.360%
24) Chloroform-d	4.339	84	265380	44.206	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.420%
26) 1,2-Dichloroethane-d4	5.024	65	162521	47.400	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.800%
32) Benzene-d6	5.043	84	559972	44.300	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.600%
36) 1,2-Dichloropropane-d6	6.063	67	179444	46.475	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.940%
41) Toluene-d8	7.307	98	515092	43.248	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.500%
43) trans-1,3-Dichloroprop...	7.612	79	83670	42.914	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.820%
47) 2-Hexanone-d5	8.082	63	168504	102.199	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.200%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	264977	48.025	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.040%
66) 1,2-Dichlorobenzene-d4	11.612	152	220046	46.198	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120822\
Data File : VV029487.D
Acq On : 08 Dec 2022 20:56
Operator : SY/MD
Sample : N5920-09
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0EJ8

Quant Time: Dec 09 04:15:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 08 01:48:46 2022
Response via : Initial Calibration

