

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120822\
 Data File : VW029482.D
 Acq On : 08 Dec 2022 18:58
 Operator : SY/MD
 Sample : N5920-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0EJ6

Quant Time: Dec 09 02:02:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 08 01:48:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	529710	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	492220	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	255332	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	108201	33.357	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	66.720%	
7) Chloroethane-d5	1.564	69	100845	40.771	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	81.540%	
11) 1,1-Dichloroethene-d2	2.105	63	149489	29.547	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	59.100%#	
21) 2-Butanone-d5	3.867	46	198414	100.272	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	100.270%	
24) Chloroform-d	4.339	84	266506	43.660	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	87.320%	
26) 1,2-Dichloroethane-d4	5.024	65	159828	45.845	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	91.700%	
32) Benzene-d6	5.043	84	559726	43.036	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	86.080%	
36) 1,2-Dichloropropane-d6	6.063	67	176780	44.498	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	89.000%	
41) Toluene-d8	7.307	98	510035	41.620	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	83.240%	
43) trans-1,3-Dichloroprop...	7.612	79	83394	41.570	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	83.140%	
47) 2-Hexanone-d5	8.082	63	163968	96.653	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	96.650%	
56) 1,1,2,2-Tetrachloroeth...	10.207	84	260916	45.960	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	91.920%	
66) 1,2-Dichlorobenzene-d4	11.612	152	209074	43.015	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	86.020%	
Target Compounds						
13) Acetone	2.159	43	5195	4.030	ug/L	Qvalue 98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120822\
Data File : VV029482.D
Acq On : 08 Dec 2022 18:58
Operator : SY/MD
Sample : N5920-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0EJ6

Quant Time: Dec 09 02:02:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 08 01:48:46 2022
Response via : Initial Calibration





#13
 Acetone
 Concen: 4.030 ug/L
 RT: 2.159 min Scan# 34
 Delta R.T. 0.003 min
 Lab File: VV029482.D
 Acq: 08 Dec 2022 18:58

Instrument :
 MSVOA_V
 ClientSampleId :
 C0EJ6

Tgt Ion: 43 Resp: 5195
 Ion Ratio Lower Upper
 43 100
 58 38.4 0.0 74.8

