

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120823\
 Data File : VW033193.D
 Acq On : 08 Dec 2023 16:15
 Operator : SY/MD
 Sample : 05685-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Dec 08 22:02:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 08 21:51:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	272874	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	269502	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	140385	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	65754	40.309	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.620%
7) Chloroethane-d5	1.532	69	51198	45.045	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.100%
11) 1,1-Dichloroethene-d2	2.060	65	27671	44.590	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.180%
21) 2-Butanone-d5	3.780	46	77302	95.410	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.410%
24) Chloroform-d	4.249	84	158660	45.832	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.660%
26) 1,2-Dichloroethane-d4	4.941	65	106221	49.352	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.700%
32) Benzene-d6	4.960	84	296860	47.098	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.200%
36) 1,2-Dichloropropane-d6	5.989	67	78724	46.452	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.900%
41) Toluene-d8	7.243	98	279963	46.562	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.120%
43) trans-1,3-Dichloroprop...	7.555	79	47477	44.302	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.600%
47) 2-Hexanone-d5	8.024	63	66370	97.483	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.480%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	113648	47.174	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.340%
66) 1,2-Dichlorobenzene-d4	11.561	152	131184	52.520	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.040%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120823\
Data File : VV033193.D
Acq On : 08 Dec 2023 16:15
Operator : SY/MD
Sample : 05685-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Dec 08 22:02:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 08 21:51:12 2023
Response via : Initial Calibration

