

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120825\
 Data File : VV039488.D
 Acq On : 08 Dec 2025 12:02
 Operator : SY/MD
 Sample : Q3787-21
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 TB01-120425

Quant Time: Dec 09 01:27:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Dec 04 00:58:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Difluorobenzene	5.566	114	6543	0.500 ug/L	0.00
5) Chlorobenzene-d5	8.785	117	6120	0.500 ug/L	0.00
12) 1,4-Dichlorobenzene-d4	11.197	152	3130	0.500 ug/L	0.02

System Monitoring Compounds					
2) Vinyl Chloride-d3	1.294	65	3588	0.545 ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 110.000%
4) 1,2-Dichloroethane-d4	4.970	65	1891	0.434 ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 86.000%
7) 1,2-Dichloropropane-d6	5.996	67	2168	0.420 ug/L	-0.02
Spiked Amount	0.500	Range	60 - 140	Recovery	= 84.000%
8) Toluene-d8	7.254	98	4063	0.466 ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	= 94.000%
10) 1,1,2,2-Tetrachloroeth...	10.133	84	1537	0.528 ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 106.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120825\
Data File : VV039488.D
Acq On : 08 Dec 2025 12:02
Operator : SY/MD
Sample : Q3787-21
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
TB01-120425

Quant Time: Dec 09 01:27:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Dec 04 00:58:46 2025
Response via : Initial Calibration

