

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV120919\
 Data File : VV013971.D
 Acq On : 09 Dec 2019 19:01
 Operator : SY/MD
 Sample : K6200-03
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFF41

Quant Time: Dec 10 06:32:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 10 04:29:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	611971	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	551210	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	240283	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	164542	46.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.98%
7) Chloroethane-d5	1.56	69	180901	48.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.94%
11) 1,1-Dichloroethene-d2	2.12	63	285729	36.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.02%
21) 2-Butanone-d5	3.92	46	234036	93.94	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.94%
24) Chloroform-d	4.39	84	335565	45.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.30%
26) 1,2-Dichloroethane-d4	5.07	65	211109	44.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.64%
32) Benzene-d6	5.09	84	680971	47.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.36%
36) 1,2-Dichloropropane-d6	6.11	67	211961	49.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.88%
41) Toluene-d8	7.35	98	589147	45.25	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.50%
43) trans-1,3-Dichloropropene-	7.66	79	89073	42.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.40%
47) 2-Hexanone-d5	8.13	63	147533	92.16	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.16%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	273556	44.06	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.12%
64) 1,2-Dichlorobenzene-d4	11.67	152	229763	48.69	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.38%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV120919\
Data File : VV013971.D
Acq On : 09 Dec 2019 19:01
Operator : SY/MD
Sample : K6200-03
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFF41

Quant Time: Dec 10 06:32:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 10 04:29:32 2019
Response via : Initial Calibration

