

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120921\
 Data File : VW023873.D
 Acq On : 09 Dec 2021 19:53
 Operator : SY/MD
 Sample : M5003-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COCT8

Quant Time: Dec 10 03:26:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 02 02:08:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	98942	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	97506	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	47505	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	25029	3.082	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	61.600%
7) Chloroethane-d5	1.568	69	24220	3.794	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.800%
11) 1,1-Dichloroethene-d2	2.108	63	38762	2.708	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	54.200%#
20) 2-Butanone-d5	3.908	46	35439	36.294	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	72.580%
24) Chloroform-d	4.349	84	66353	4.691	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.800%
26) 1,2-Dichloroethane-d4	5.037	65	32465	4.913	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.200%
32) Benzene-d6	5.050	84	121090	4.559	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.200%
36) 1,2-Dichloropropane-d6	6.069	67	35845	4.814	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.200%
41) Toluene-d8	7.317	98	105176	4.238	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.800%
43) trans-1,3-Dichloroprop...	7.625	79	13738	4.577	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.600%
46) 2-Hexanone-d5	8.091	63	57703	57.862	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	115.720%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	26933	5.026	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.600%
66) 1,2-Dichlorobenzene-d4	11.625	152	45119	5.372	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120921\
Data File : VV023873.D
Acq On : 09 Dec 2021 19:53
Operator : SY/MD
Sample : M5003-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COCT8

Quant Time: Dec 10 03:26:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 02 02:08:23 2021
Response via : Initial Calibration

