

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120922\
 Data File : VW029517.D
 Acq On : 09 Dec 2022 13:36
 Operator : SY/MD
 Sample : N5920-11
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0EN1

Quant Time: Dec 09 23:50:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 09 23:48:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	467407	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	423349	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.236	152	223163	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	143804	50.242	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 100.480%			
7) Chloroethane-d5	1.561	69	112394	51.497	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 103.000%			
11) 1,1-Dichloroethene-d2	2.101	63	175677	39.352	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 78.700%			
21) 2-Butanone-d5	3.870	46	174676	100.042	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 100.040%			
24) Chloroform-d	4.339	84	250295	46.470	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 92.940%			
26) 1,2-Dichloroethane-d4	5.021	65	147631	47.991	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 95.980%			
32) Benzene-d6	5.040	84	553995	49.524	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 99.040%			
36) 1,2-Dichloropropane-d6	6.063	67	166882	48.840	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 97.680%			
41) Toluene-d8	7.307	98	493438	46.816	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 93.640%			
43) trans-1,3-Dichloroprop...	7.612	79	68538	39.723	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 79.440%			
47) 2-Hexanone-d5	8.082	63	152398	104.447	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 104.450%			
56) 1,1,2,2-Tetrachloroeth...	10.204	84	230416	47.190	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 94.380%			
66) 1,2-Dichlorobenzene-d4	11.612	152	197192	46.418	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 92.840%			
Target Compounds					Qvalue	
13) Acetone	2.162	43	2854	2.509	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120922\
Data File : VV029517.D
Acq On : 09 Dec 2022 13:36
Operator : SY/MD
Sample : N5920-11
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0EN1

Quant Time: Dec 09 23:50:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 09 23:48:53 2022
Response via : Initial Calibration

