

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120922\
 Data File : VW029546.D
 Acq On : 10 Dec 2022 01:09
 Operator : SY/MD
 Sample : N5915-11
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5N12

Quant Time: Dec 10 01:36:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 10 00:41:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	481422	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	439972	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	225859	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	136707	46.372	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.740%
7) Chloroethane-d5	1.564	69	111531	49.614	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.220%
11) 1,1-Dichloroethene-d2	2.105	63	171520	37.302	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.600%
21) 2-Butanone-d5	3.873	46	176722	98.268	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.270%
24) Chloroform-d	4.339	84	250767	45.203	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.400%
26) 1,2-Dichloroethane-d4	5.024	65	149080	47.051	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.100%
32) Benzene-d6	5.043	84	544813	46.863	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.720%
36) 1,2-Dichloropropane-d6	6.063	67	165716	46.667	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.340%
41) Toluene-d8	7.307	98	486832	44.444	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.880%
43) trans-1,3-Dichloroprop...	7.612	79	65629	36.600	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.200%
47) 2-Hexanone-d5	8.082	63	149422	98.538	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.540%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	230579	45.439	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.880%
66) 1,2-Dichlorobenzene-d4	11.612	152	195101	45.378	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120922\
Data File : VV029546.D
Acq On : 10 Dec 2022 01:09
Operator : SY/MD
Sample : N5915-11
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5N12

Quant Time: Dec 10 01:36:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 10 00:41:04 2022
Response via : Initial Calibration

