

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120922\
 Data File : VW029554.D
 Acq On : 10 Dec 2022 04:17
 Operator : SY/MD
 Sample : N5915-15
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5N13

Quant Time: Dec 10 04:44:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 10 00:41:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	477763	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	448212	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	231895	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	130865	44.731	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.460%
7) Chloroethane-d5	1.565	69	108907	48.818	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.640%
11) 1,1-Dichloroethene-d2	2.105	63	166428	36.472	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.940%
21) 2-Butanone-d5	3.873	46	177648	99.539	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.540%
24) Chloroform-d	4.339	84	243480	44.225	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.460%
26) 1,2-Dichloroethane-d4	5.024	65	147580	46.935	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.860%
32) Benzene-d6	5.044	84	526834	44.484	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.960%
36) 1,2-Dichloropropane-d6	6.063	67	161907	44.756	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.520%
41) Toluene-d8	7.307	98	472590	42.351	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.700%
43) trans-1,3-Dichloroprop...	7.613	79	63318	34.662	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.320%
47) 2-Hexanone-d5	8.082	63	149927	97.053	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.050%
56) 1,1,2,2-Tetrachloroeth...	10.204	84	226271	43.770	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.540%
66) 1,2-Dichlorobenzene-d4	11.612	152	191455	43.371	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.740%
Target Compounds						
13) Acetone	2.166	43	8076	6.946	ug/L	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120922\
Data File : VV029554.D
Acq On : 10 Dec 2022 04:17
Operator : SY/MD
Sample : N5915-15
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5N13

Quant Time: Dec 10 04:44:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 10 00:41:04 2022
Response via : Initial Calibration

