

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120922\
 Data File : VV029512.D
 Acq On : 09 Dec 2022 11:39
 Operator : SY/MD
 Sample : VV1209WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK303

Quant Time: Dec 09 23:49:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 09 23:48:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	529741	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	475243	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	251022	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	166710	51.392	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 102.780%	
7) Chloroethane-d5	1.564	69	136284	55.096	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 110.200%	
11) 1,1-Dichloroethene-d2	2.105	63	207814	41.073	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 82.140%	
21) 2-Butanone-d5	3.870	46	216338	109.324	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 109.320%	
24) Chloroform-d	4.339	84	297534	48.741	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 97.480%	
26) 1,2-Dichloroethane-d4	5.024	65	177260	50.843	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 101.680%	
32) Benzene-d6	5.043	84	664946	52.952	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 105.900%	
36) 1,2-Dichloropropane-d6	6.063	67	202195	52.713	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 105.420%	
41) Toluene-d8	7.307	98	598130	50.552	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 101.100%	
43) trans-1,3-Dichloroprop...	7.612	79	82144	42.410	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 84.820%	
47) 2-Hexanone-d5	8.082	63	187015	114.176	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 114.180%	
56) 1,1,2,2-Tetrachloroeth...	10.207	84	273276	49.856	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 99.720%	
66) 1,2-Dichlorobenzene-d4	11.612	152	238271	49.863	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 99.720%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120922\
Data File : VV029512.D
Acq On : 09 Dec 2022 11:39
Operator : SY/MD
Sample : VV1209WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK303

Quant Time: Dec 09 23:49:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 09 23:48:53 2022
Response via : Initial Calibration

