

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120922\
 Data File : VW029523.D
 Acq On : 09 Dec 2022 15:57
 Operator : SY/MD
 Sample : N5920-17
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0EP0

Quant Time: Dec 09 23:52:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 09 23:48:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	443727	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	408665	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	208686	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	144616	53.222	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	= 106.440%		
7) Chloroethane-d5	1.564	69	118996	57.432	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	= 114.860%		
11) 1,1-Dichloroethene-d2	2.105	63	184513	43.537	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	= 87.080%		
21) 2-Butanone-d5	3.867	46	183700	110.825	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	= 110.830%		
24) Chloroform-d	4.339	84	263436	51.520	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	= 103.040%		
26) 1,2-Dichloroethane-d4	5.024	65	157560	53.952	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	= 107.900%		
32) Benzene-d6	5.043	84	581420	53.844	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	= 107.680%		
36) 1,2-Dichloropropane-d6	6.063	67	174788	52.992	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	= 105.980%		
41) Toluene-d8	7.307	98	517548	50.868	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	= 101.740%		
43) trans-1,3-Dichloroprop...	7.612	79	72267	43.389	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	= 86.780%		
47) 2-Hexanone-d5	8.082	63	155638	110.500	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	= 110.500%		
56) 1,1,2,2-Tetrachloroeth...	10.204	84	244496	51.873	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	= 103.740%		
66) 1,2-Dichlorobenzene-d4	11.612	152	202994	51.099	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	= 102.200%		
Target Compounds						
13) Acetone	2.159	43	2751	2.548	ug/L	Qvalue 97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120922\
Data File : VV029523.D
Acq On : 09 Dec 2022 15:57
Operator : SY/MD
Sample : N5920-17
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0EP0

Quant Time: Dec 09 23:52:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 09 23:48:53 2022
Response via : Initial Calibration





#13
 Acetone
 Concen: 2.548 ug/L
 RT: 2.159 min Scan# 34
 Delta R.T. 0.003 min
 Lab File: VV029523.D
 Acq: 09 Dec 2022 15:57

Instrument :
 MSVOA_V
 ClientSampleId :
 C0EP0

Tgt Ion: 43 Resp: 2751
 Ion Ratio Lower Upper
 43 100
 58 39.2 0.0 74.8

