

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120925\
 Data File : VV039506.D
 Acq On : 09 Dec 2025 16:03
 Operator : SY/MD
 Sample : Q3787-08
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 OWBR-02-170-120425-SIM

Quant Time: Dec 10 01:17:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 09 08:40:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	6146	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	6167	0.500	ug/L	0.02
12) 1,4-Dichlorobenzene-d4	11.175	152	3302	0.500	ug/L	0.00

System Monitoring Compounds						
2) Vinyl Chloride-d3	1.295	65	3646	0.604	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	120.000%
4) 1,2-Dichloroethane-d4	4.994	65	1954	0.477	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	96.000%
7) 1,2-Dichloropropane-d6	5.996	67	2259	0.434	ug/L	-0.02
Spiked Amount	0.500	Range	60 - 140	Recovery	=	86.000%
8) Toluene-d8	7.254	98	4324	0.493	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	98.000%
10) 1,1,2,2-Tetrachloroeth...	10.133	84	1687	0.575	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	116.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120925\
Data File : VV039506.D
Acq On : 09 Dec 2025 16:03
Operator : SY/MD
Sample : Q3787-08
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
OWBR-02-170-120425-SIM

Quant Time: Dec 10 01:17:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 09 08:40:48 2025
Response via : Initial Calibration

