

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121019\
 Data File : VV013993.D
 Acq On : 10 Dec 2019 17:48
 Operator : SY/MD
 Sample : K6202-12
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0C09

Quant Time: Dec 11 02:30:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 11 00:25:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	625152	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	544765	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	241931	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	145220	40.16	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.32%
7) Chloroethane-d5	1.56	69	157874	41.83	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.66%
11) 1,1-Dichloroethene-d2	2.12	63	246803	30.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.74%
21) 2-Butanone-d5	3.92	46	263042	103.36	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.36%
24) Chloroform-d	4.39	84	360069	47.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.84%
26) 1,2-Dichloroethane-d4	5.07	65	227381	47.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.52%
32) Benzene-d6	5.09	84	720755	51.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.12%
36) 1,2-Dichloropropane-d6	6.11	67	189230	44.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.32%
41) Toluene-d8	7.35	98	497680	38.67	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	77.34%#
43) trans-1,3-Dichloropropene-	7.66	79	78635	38.14	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.28%
47) 2-Hexanone-d5	8.13	63	125789	79.50	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.50%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	241028	39.28	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	78.56%
64) 1,2-Dichlorobenzene-d4	11.67	152	220093	46.32	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.64%

Target Compounds					Ovalue
13) Acetone	2.18	43	7782	2.087 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121019\
Data File : VV013993.D
Acq On : 10 Dec 2019 17:48
Operator : SY/MD
Sample : K6202-12
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0C09

Quant Time: Dec 11 02:30:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 11 00:25:42 2019
Response via : Initial Calibration

