

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121019\
 Data File : VV013998.D
 Acq On : 10 Dec 2019 19:50
 Operator : SY/MD
 Sample : K6200-06
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFF24

Quant Time: Dec 11 03:04:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 11 00:25:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	580373	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	566932	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	236611	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	142324	42.40	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.80%
7) Chloroethane-d5	1.57	69	163435	46.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.30%
11) 1,1-Dichloroethene-d2	2.13	63	243480	32.80	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.60%
21) 2-Butanone-d5	3.92	46	241186	102.09	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.09%
24) Chloroform-d	4.40	84	307135	43.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.14%
26) 1,2-Dichloroethane-d4	5.08	65	197092	44.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.24%
32) Benzene-d6	5.09	84	610012	41.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.04%
36) 1,2-Dichloropropane-d6	6.11	67	180825	41.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	82.02%
41) Toluene-d8	7.36	98	510079	38.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	76.18%#
43) trans-1,3-Dichloropropene-	7.66	79	79642	37.12	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.24%
47) 2-Hexanone-d5	8.13	63	132938	80.74	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	80.74%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	260441	40.79	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.58%
64) 1,2-Dichlorobenzene-d4	11.67	152	212697	45.77	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.54%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121019\
Data File : VV013998.D
Acq On : 10 Dec 2019 19:50
Operator : SY/MD
Sample : K6200-06
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFF24

Quant Time: Dec 11 03:04:01 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 11 00:25:42 2019
Response via : Initial Calibration

