

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121020\
 Data File : VV019535.D
 Acq On : 11 Dec 2020 01:27
 Operator : SY/MD
 Sample : L4997-01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AN5

Quant Time: Dec 11 04:35:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 11 02:38:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	673663	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	627575	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.25	152	291820	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	195845	40.54	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.08%
7) Chloroethane-d5	1.57	69	169024	44.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.72%
11) 1,1-Dichloroethene-d2	2.11	63	301021	33.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.14%
21) 2-Butanone-d5	3.89	46	249891	95.17	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.17%
24) Chloroform-d	4.35	84	396746	45.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.14%
26) 1,2-Dichloroethane-d4	5.04	65	277724	48.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.62%
32) Benzene-d6	5.05	84	787880	47.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.14%
36) 1,2-Dichloropropane-d6	6.07	67	249526	47.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.18%
41) Toluene-d8	7.32	98	704140	46.72	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.44%
43) trans-1,3-Dichloropropene-	7.62	79	132956	44.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.16%
47) 2-Hexanone-d5	8.09	63	192913	93.75	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.75%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	328977	46.39	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.78%
64) 1,2-Dichlorobenzene-d4	11.63	152	261573	47.95	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.90%

Target Compounds					Ovalue
34) Trichloroethene	5.92	95	31669	6.536 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121020\
Data File : VV019535.D
Acq On : 11 Dec 2020 01:27
Operator : SY/MD
Sample : L4997-01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AN5

Quant Time: Dec 11 04:35:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 11 02:38:22 2020
Response via : Initial Calibration

