

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121021\
Data File : VV023893.D
Acq On : 10 Dec 2021 17:51
Operator : SY/MD
Sample : M5003-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0CW7




Manual IntegrationsAPPROVED

Quant Time: Dec 11 04:46:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Dec 11 04:33:17 2021
Response via : Initial Calibration

Reviewed By :Mahesh
Dadoda

12/13/2021
Supervised By :Semsettin
Yesilyurt

12/13/2021



TIC: VV023893.D\data.ms

(20) 2-Butanone-d5 (S)

3.912min (+ 0.010) 37.82 ug/L

response 41481

Ion	Exp%	Act%
46.10	100.00	100.00
77.10	9.40	28.01#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121021\
 Data File : VV023893.D
 Acq On : 10 Dec 2021 17:51
 Operator : SY/MD
 Sample : M5003-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1




Instrument :
 MSVOA_V
 ClientSampleId :
 C0CW7

Manual IntegrationsAPPROVED

Quant Time: Dec 11 04:46:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Dec 11 04:33:17 2021
 Response via : Initial Calibration

Reviewed By :Mahesh
 Dadoda

12/13/2021
 Supervised By :Semsettin
 Yesilyurt



TIC: VV023893.D\data.ms

(20) 2-Butanone-d5 (S)

3.912min (+ 0.010) 56.78 ug/L m

response 62281

Ion	Exp%	Act%
46.10	100.00	100.00
77.10	9.40	18.65#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121021\
 Data File : VW023893.D
 Acq On : 10 Dec 2021 17:51
 Operator : SY/MD
 Sample : M5003-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0CW7

Manual IntegrationsAPPROVED

Reviewed By :Mahesh
 Dadoda

12/13/2021
 Supervised By :Semsettin
 Yesilyurt

Quant Time: Dec 11 04:46:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Dec 11 04:33:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	104441	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	109568	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	54644	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	25982	4.628	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	92.600%	
7) Chloroethane-d5	1.568	69	23667	5.052	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	101.000%	
11) 1,1-Dichloroethene-d2	2.108	63	35211	3.213	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	64.200%	
20) 2-Butanone-d5	3.912	46	62281m	56.777	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	113.560%	
24) Chloroform-d	4.352	84	62075	4.881	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.600%	
26) 1,2-Dichloroethane-d4	5.037	65	30497	5.150	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	103.000%	
32) Benzene-d6	5.056	84	104410	4.811	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.200%	
36) 1,2-Dichloropropane-d6	6.072	67	33100	5.055	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	101.200%	
41) Toluene-d8	7.320	98	85952	4.399	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	88.000%	
43) trans-1,3-Dichloroprop...	7.628	79	12173	4.482	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	89.600%	
46) 2-Hexanone-d5	8.091	63	58558	48.159	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	96.320%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	26070	4.800	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	96.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	39927	5.470	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	109.400%	
Target Compounds						
					Qvalue	
34) Trichloroethene	5.924	95	5819	0.666	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121021\
Data File : VV023893.D
Acq On : 10 Dec 2021 17:51
Operator : SY/MD
Sample : M5003-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0CW7

Manual IntegrationsAPPROVED

Quant Time: Dec 11 04:46:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Dec 11 04:33:17 2021
Response via : Initial Calibration

Reviewed By :Mahesh
Dadoda

12/13/2021
Supervised By :Semsettin
Yesilyurt

12/13/2021

