

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121119\
 Data File : VV014013.D
 Acq On : 11 Dec 2019 15:27
 Operator : SY/MD
 Sample : K6308-09
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0C26

Quant Time: Dec 12 03:39:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 12 03:01:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	606467	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	597803	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	245051	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	184294	52.54	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	105.08%
7) Chloroethane-d5	1.56	69	191241	52.24	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	104.48%
11) 1,1-Dichloroethene-d2	2.12	63	279672	36.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.12%
21) 2-Butanone-d5	3.92	46	209546	84.88	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	84.88%
24) Chloroform-d	4.39	84	332153	45.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.18%
26) 1,2-Dichloroethane-d4	5.07	65	211322	45.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.54%
32) Benzene-d6	5.09	84	690837	44.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.20%
36) 1,2-Dichloropropane-d6	6.11	67	215565	46.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.72%
41) Toluene-d8	7.35	98	593853	42.05	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.10%
43) trans-1,3-Dichloropropene-	7.66	79	88352	39.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.10%
47) 2-Hexanone-d5	8.13	63	117831	67.87	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	67.87%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	271144	40.27	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	80.54%
64) 1,2-Dichlorobenzene-d4	11.67	152	242523	50.39	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.78%

Target Compounds					Ovalue
13) Acetone	2.18	43	3559	0.984 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121119\
Data File : VV014013.D
Acq On : 11 Dec 2019 15:27
Operator : SY/MD
Sample : K6308-09
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0C26

Quant Time: Dec 12 03:39:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 12 03:01:11 2019
Response via : Initial Calibration

