

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121123\
 Data File : VW033217.D
 Acq On : 11 Dec 2023 16:54
 Operator : SY/MD
 Sample : 05685-19
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Dec 12 00:42:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 08 21:51:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	263839	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	266169	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	138230	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	59475	37.709	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	75.420%		
7) Chloroethane-d5	1.532	69	49134	44.710	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	89.420%		
11) 1,1-Dichloroethene-d2	2.060	65	25347	42.244	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	84.480%		
21) 2-Butanone-d5	3.780	46	80142	102.303	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	102.300%		
24) Chloroform-d	4.249	84	160811	48.044	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	96.080%		
26) 1,2-Dichloroethane-d4	4.940	65	106315	51.087	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	102.180%		
32) Benzene-d6	4.960	84	294198	47.260	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.520%		
36) 1,2-Dichloropropane-d6	5.989	67	79635	47.578	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	95.160%		
41) Toluene-d8	7.243	98	263325	44.343	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	88.680%		
43) trans-1,3-Dichloroprop...	7.551	79	45561	43.047	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	86.100%		
47) 2-Hexanone-d5	8.024	63	65283	97.087	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	97.090%		
56) 1,1,2,2-Tetrachloroeth...	10.152	84	118895	49.970	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	99.940%		
66) 1,2-Dichlorobenzene-d4	11.561	152	128828	52.381	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	104.760%		
Target Compounds						Qvalue
8) Chloroethane	1.281	64	885	0.911	ug/L #	1
13) Acetone	2.111	43	26177	35.003	ug/L	95
27) 1,2-Dichloroethane	5.066	62	3538	1.354	ug/L #	94
44) trans-1,3-Dichloropropene	7.551	75	2295	0.801	ug/L	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121123\
Data File : VV033217.D
Acq On : 11 Dec 2023 16:54
Operator : SY/MD
Sample : 05685-19
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Dec 12 00:42:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 08 21:51:12 2023
Response via : Initial Calibration









#8
Chloroethane
Concen: 0.911 ug/L
RT: 1.281 min Scan# 71
Delta R.T. -0.267 min
Lab File: VV033217.D
Acq: 11 Dec 2023 16:54




Instrument :
MSVOA_V
ClientSampleId :


Tgt Ion: 64 Resp: 885
Ion Ratio Lower Upper
64 100
66 147.7 22.3 41.5#



#13
Acetone
Concen: 35.003 ug/L
RT: 2.111 min Scan# 333
Delta R.T. 0.003 min
Lab File: VV033217.D
Acq: 11 Dec 2023 16:54

Tgt Ion: 43 Resp: 26177
Ion Ratio Lower Upper
43 100
58 34.3 0.0 62.8





#27

1,2-Dichloroethane

Concen: 1.354 ug/L

RT: 5.066 min Scan# 1

Delta R.T. 0.029 min

Lab File: VV033217.D

Acq: 11 Dec 2023 16:54

Instrument :

MSVOA_V





ClientSampleId :

Tgt Ion: 62 Resp: 3538

Ion Ratio Lower Upper

62 100

98 12.0 7.8 11.8#



#44

trans-1,3-Dichloropropene

Concen: 0.801 ug/L

RT: 7.551 min Scan# 2025

Delta R.T. -0.029 min

Lab File: VV033217.D

Acq: 11 Dec 2023 16:54

Tgt Ion: 75 Resp: 2295

Ion Ratio Lower Upper

75 100

77 37.6 22.0 41.0

