

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121123\
 Data File : VW033215.D
 Acq On : 11 Dec 2023 16:08
 Operator : SY/MD
 Sample : 05685-17
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YCH82

Quant Time: Dec 12 00:42:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 08 21:51:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	181191	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	184099	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	92965	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	40832	37.697	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.400%
7) Chloroethane-d5	1.529	69	33794	44.778	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.560%
11) 1,1-Dichloroethene-d2	2.060	65	17707	42.972	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.940%
21) 2-Butanone-d5	3.780	46	74938	139.294	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	139.290%#
24) Chloroform-d	4.249	84	114143	49.656	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.320%
26) 1,2-Dichloroethane-d4	4.941	65	76581	53.585	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.160%
32) Benzene-d6	4.960	84	206032	47.851	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.700%
36) 1,2-Dichloropropane-d6	5.989	67	57788	49.917	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.840%
41) Toluene-d8	7.243	98	183970	44.790	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.580%
43) trans-1,3-Dichloroprop...	7.555	79	32031	43.755	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.500%
47) 2-Hexanone-d5	8.021	63	58382	125.530	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	125.530%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	95029	57.744	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	115.480%
66) 1,2-Dichlorobenzene-d4	11.561	152	89837	54.313	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.620%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121123\
Data File : VV033215.D
Acq On : 11 Dec 2023 16:08
Operator : SY/MD
Sample : 05685-17
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YCH82

Quant Time: Dec 12 00:42:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 08 21:51:12 2023
Response via : Initial Calibration

