

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121123\
 Data File : VW033220.D
 Acq On : 11 Dec 2023 18:04
 Operator : SY/MD
 Sample : 05730-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0GA6

Quant Time: Dec 12 00:43:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 08 21:51:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	262741	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	259027	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	136545	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	58442	37.208	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	74.420%
7) Chloroethane-d5	1.532	69	47388	43.301	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.600%
11) 1,1-Dichloroethene-d2	2.060	65	24351	40.754	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.500%
21) 2-Butanone-d5	3.780	46	80585	103.298	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.300%
24) Chloroform-d	4.249	84	153639	46.093	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.180%
26) 1,2-Dichloroethane-d4	4.941	65	100223	48.361	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.720%
32) Benzene-d6	4.960	84	276776	45.687	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.380%
36) 1,2-Dichloropropane-d6	5.989	67	76451	46.935	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.880%
41) Toluene-d8	7.243	98	251503	43.520	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.040%
43) trans-1,3-Dichloroprop...	7.551	79	44261	42.972	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.940%
47) 2-Hexanone-d5	8.021	63	63931	97.698	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.700%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	117730	50.844	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.680%
66) 1,2-Dichlorobenzene-d4	11.561	152	124816	51.376	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121123\
Data File : VV033220.D
Acq On : 11 Dec 2023 18:04
Operator : SY/MD
Sample : 05730-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0GA6

Quant Time: Dec 12 00:43:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 08 21:51:12 2023
Response via : Initial Calibration

